

DATASHEET · ANESTHESIA & VENTILATION

Anesthesia Induction Chamber.

SKU CS-ANES-CHAMBER-001 Spec v8 Rev. 2026-05-26

DATASHEET · P
CS-ANES-CHAMBE
Rev 2C



PRODUCT FAMILY	Small-animal inhalation induction chamber
ANIMAL MODEL	Mouse, rat, rabbit, cat (model-dependent)
APPLICATION	Pre-procedure inhalation anesthesia induction
COUNTRY OF ORIGIN	China
MFG. STANDARD	ISO 9001 supplier
LOT TRACEABLE	Yes · per unit
VERIFIED	Spec v8 · May 2026

CHAMBER-LEVEL SPECIFICATIONS

BODY, LID, CLOSURE, GAS

BODY MATERIAL	Cast acrylic (PMMA), transparent	LID THICKNESS	11 mm, hinged
CLOSURE	Seal-strip, compression	GAS INTERFACE	Coaxial or separate inlet/outlet
EXHAUST ROUTING	Corrugated tubing charcoal canister	CONFIGURATIONS	7 model variants (4 coaxial, 3 sep I/O)
CLEANING	Wipe-clean, acrylic-safe disinfectant	WARRANTY	12 months from ship date

MODEL VARIANTS · 7 LINE ITEMS

EXTENDED ATTRIBUTES

QTY	MODEL	SKU	I/O	MATERIAL	EXTER
1	Mouse, coaxial I/O	CS-ANES-CHAMBER-MOUSE-COAX	Coaxial	PMMA	150 ×
1	Mouse and rat, coaxial I/O	CS-ANES-CHAMBER-RODENT-COAX	Coaxial	PMMA	240 ×
1	Rabbit and cat, coaxial I/O	CS-ANES-CHAMBER-RABCAT-COAX	Coaxial	PMMA	380 ×
1	Large animal, coaxial I/O	CS-ANES-CHAMBER-LARGE-COAX	Coaxial	PMMA	500 ×
1	Mouse, separate I/O	CS-ANES-CHAMBER-MOUSE-SEP	Separate	PMMA	150 ×
1	Mouse and rat, separate I/O	CS-ANES-CHAMBER-RODENT-SEP	Separate	PMMA	240 ×
1	Rabbit and cat, separate I/O	CS-ANES-CHAMBER-RABCAT-SEP	Separate	PMMA	380 ×

Scope. Chamber scope excludes the anesthetic delivery system, waste-gas scavenging, warming, and monitoring. Pair this chamber with separately
equipment sized to your animal model and approved animal-care protocol. See the accompanying Setup Guide PDF for staging.

INFO@CONDUCTSCIENCE.COM CONDUCTSCIENCE.COM

CS-ANES-CHAMBER-001 · DATASH

DATASHEET · ANESTHESIA & VENTILATION

Anesthesia Induction Chamber.

SKU CS-ANES-CHAMBER-001 Spec v8 Rev. 2026-05-26

DATASHEET · PRODUCT
CS-ANES-CHAMBER
Rev 2026-05-26

MODEL VARIANTS · EXTENDED ATTRIBUTES

SAME MODEL ORDER

ITEM	ANIMAL FIT	INTERNAL FLOOR (APPROX)	WALL	PORT FITTING
Mouse, coaxial I/O	Mouse	138 × 88	6 mm	Barbed, 6 mm
Mouse and rat, coaxial I/O	Mouse, rat	228 × 108	6 mm	Barbed, 6 mm
Rabbit and cat, coaxial I/O	Rabbit, cat	366 × 266	8 mm	Barbed, 10 mm
Large animal, coaxial I/O	Large rodent / small lagomorph	486 × 286	8 mm	Barbed, 10 mm
Mouse, separate I/O	Mouse	138 × 88	6 mm	Barbed, 6 mm
Mouse and rat, separate I/O	Mouse, rat	228 × 108	6 mm	Barbed, 6 mm
Rabbit and cat, separate I/O	Rabbit, cat	366 × 266	8 mm	Barbed, 10 mm

MATERIALS REFERENCE

4 MATERIALS

CODE	DESCRIPTION	PROPERTIES	USED IN
PMMA	Cast acrylic, transparent	Light transmission >92%	Chamber body, lid
SS 304	Stainless steel	Corrosion-resistant	Hinges, seal clasps
Silicone	Seal strip	Compression gasket	Lid closure
Brass	Nickel-plated fitting	Barbed gas port	Inlet / outlet ports

Cross-reference. Internal floor dimensions are approximate, derived from external dimensions and wall thickness. Confirm chamber fit, gas-fitting compatibility, and scavenging before ordering. Sterilization guidance reflects cast-acrylic compatibility — do not autoclave.

DATASHEET · ANESTHESIA & VENTILATION

Anesthesia Induction Chamber.

SKU CS-ANES-CHAMBER-001 Spec v8 Rev. 2026-05-26

DATASHEET · P
CS-ANES-CHAMBE
Rev 20

OPERATING & STERILIZATION ENVELOPES

CONDITIONS, CYCLES, COM

OPERATING		STORAGE	
TEMPERATURE	15–30 °C ambient	TEMPERATURE	5–40 °C
HUMIDITY	20–80% RH, non-condensing	HUMIDITY	<80% RH
PRESSURE	Ambient, low-flow gas	ORIENTATION	Flat, lid closed
STERILIZATION & CLEANING			
AUTOCLAVE	No — crazes PMMA	ETO	Compatible (validate)
GAMMA	No — yellows PMMA	WIPE	Mild detergent, 70% IPA wipe
AVOID	Acetone, aromatic / chlorinated solvents		

ANIMAL COMPATIBILITY BY SPECIES

SPECIES	BODY WEIGHT	CHAMBER FIT
Mouse	15–40 g	150 / 240 mm chambers
Rat	150–500 g	240 mm chamber
Rabbit	2–5 kg	380 mm chamber
Cat	3–6 kg	380 / 500 mm chambers

NOTES USE & S

- Research use only — not a medical device
- Pair with vaporizer, scavenging
- Use under approved IACUC protocol
- Waste-gas scavengi required

PMMA COAXIAL / SEPARATE I/O WIPE-CLEAN
RESEARCH USE ONLY

Specifications verified against current published product data. Verified May 2026 · Since May 2026.

Cite: ConductScience CS-ANES-CHAMBER-001, Spec v8, May 2026. conductscience.com/lab/anesthesia-induction-chamber

INFO@CONDUCTSCIENCE.COM CONDUCTSCIENCE.COM

CS-ANES-CHAMBER-001 · DATASH